

SUPERVISOR TEST

UTICA CITY  
 POLICE DEPT  
 DATAMASTER dmt: 107308

*RMS*  
*21-50843*

Date: 12/29/2021

OPERATOR NAME:  
 MICHAEL J CLIVE  
 Permit #: 36682  
 Operator Agency/Dept: UTICA CITY PD  
 Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:07
INTERNAL STANDARD	VERIFIED	15:07
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	15:07
BLANK TEST	0.00	15:08
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	15:08
BLANK TEST	0.00	15:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	15:09
BLANK TEST	0.00	15:10

