

SUPERVISOR TEST

UTICA CITY  
 POLICE DEPT  
 DATAMASTER dmt: 100282

*RMS*  
*23-12291*

Date: 03/21/2023

OPERATOR NAME:  
 MICHAEL J CLIVE  
 Permit #: 36682  
 Operator Agency/Dept: UTICA CITY PD  
 Reference Standard #: 22420

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:31
INTERNAL STANDARD	VERIFIED	13:32
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:32
BLANK TEST	0.00	13:33
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:33
BLANK TEST	0.00	13:34
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:34
BLANK TEST	0.00	13:35

