

SUPERVISOR TEST

UTICA CITY  
 POLICE DEPT  
 DATAMASTER dmt: 100282

Date: 05/30/2023

*RMS*  
*23-21/230*

OPERATOR NAME:  
 MICHAEL J CLIVE  
 Permit #: 36682  
 Operator Agency/Dept: UTICA CITY PD  
 Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:56
INTERNAL STANDARD	VERIFIED	11:57
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	11:57
BLANK TEST	0.00	11:58
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	11:58
BLANK TEST	0.00	11:59
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	11:59
BLANK TEST	0.00	12:00

