

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 06/06/2023

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:54
INTERNAL STANDARD	VERIFIED	09:54
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	09:54
BLANK TEST	0.00	09:55
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	09:55
BLANK TEST	0.00	09:56
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	09:56
BLANK TEST	0.00	09:57

