

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 11/22/2023

RMJ
23-52491

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:55
INTERNAL STANDARD	VERIFIED	12:55
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:55
BLANK TEST	0.00	12:56
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:56
BLANK TEST	0.00	12:57
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:58
BLANK TEST	0.00	12:58

