

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 12/26/2023

RMJ
23-57328

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:18
INTERNAL STANDARD	VERIFIED	12:18
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:18
BLANK TEST	0.00	12:19
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:20
BLANK TEST	0.00	12:20
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:21
BLANK TEST	0.00	12:22

