

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 01/31/2023

RMS
23-4719

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:37
INTERNAL STANDARD	VERIFIED	12:37
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:37
BLANK TEST	0.00	12:38
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:38
BLANK TEST	0.00	12:39
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:40
BLANK TEST	0.00	12:40

