

SUPERVISOR TEST

UTICA CITY  
 POLICE DEPT  
 DATAMASTER dmt: 107308

Date: 04/18/2023

OPERATOR NAME:  
 MICHAEL J CLIVE  
 Permit #: 36682  
 Operator Agency/Dept: UTICA CITY PD  
 Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:18
INTERNAL STANDARD	VERIFIED	11:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	11:18
BLANK TEST	0.00	11:19
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	11:20
BLANK TEST	0.00	11:20
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	11:21
BLANK TEST	0.00	11:22

*Work Station*

*MS  
 23-16978*

