

SUPERVISOR TEST

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 107308

Date: 06/27/2023

*RMJ*  
*23-28922*

OPERATOR NAME:  
MICHAEL J CLIVE  
Permit #: 36682  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:52
INTERNAL STANDARD	VERIFIED	10:52
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	10:52
BLANK TEST	0.00	10:53
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	10:53
BLANK TEST	0.00	10:54
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	10:55
BLANK TEST	0.00	10:55

