

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 107308

Date: 09/12/2023

RMS
23-11713

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:54
INTERNAL STANDARD	VERIFIED	10:54
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	10:54
BLANK TEST	0.00	10:55
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	10:55
BLANK TEST	0.00	10:56
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	10:56
BLANK TEST	0.00	10:57

