

SUPERVISOR TEST

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 107308

*LM*  
*23-48293*

Date: 10/25/2023

OPERATOR NAME:  
MICHAEL J CLIVE  
Permit #: 36682  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:02
INTERNAL STANDARD	VERIFIED	09:02
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	09:02
BLANK TEST	0.00	09:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	09:03
BLANK TEST	0.00	09:04
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	09:04
BLANK TEST	0.00	09:05

