

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 03/23/2021

*RMS
21-10257*

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:08
INTERNAL STANDARD	VERIFIED	13:08
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:08
BLANK TEST	0.00	13:09
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:09
BLANK TEST	0.00	13:10
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:10
BLANK TEST	0.00	13:11

