

SUPERVISOR TEST

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 100282

Date: 05/04/2021

OPERATOR NAME:  
MICHAEL J CLIVE  
Permit #: 36682  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 21020

*RMS*  
*21-16597*

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:32
INTERNAL STANDARD	VERIFIED	15:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:33
BLANK TEST	0.00	15:33
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	15:34
BLANK TEST	0.00	15:35
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	15:35
BLANK TEST	0.00	15:36

