

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 100282

RMS

Date: 10/12/2021

71-1/0/84

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:10
INTERNAL STANDARD	VERIFIED	13:11
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:11
BLANK TEST	0.00	13:12
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:12
BLANK TEST	0.00	13:13
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:13
BLANK TEST	0.00	13:14

