

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 10/26/2021

ZMS
21-42497

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:54
INTERNAL STANDARD	VERIFIED	12:54
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:54
BLANK TEST	0.00	12:55
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:55
BLANK TEST	0.00	12:56
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:56
BLANK TEST	0.00	12:57

