

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 04/14/2021

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 21020

RMS
21-17479

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:12
INTERNAL STANDARD	VERIFIED	12:12
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:12
BLANK TEST	0.00	12:13
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:13
BLANK TEST	0.00	12:14
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:14
BLANK TEST	0.00	12:15

