

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 06/08/2021

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 21100

RMS
21-22073

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:07
INTERNAL STANDARD	VERIFIED	13:07
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	13:08
BLANK TEST	0.00	13:08
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	13:09
BLANK TEST	0.00	13:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	13:10
BLANK TEST	0.00	13:11

