

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 08/10/2021

2m)
21-31791

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:53
INTERNAL STANDARD	VERIFIED	12:53
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:53
BLANK TEST	0.00	12:54
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:54
BLANK TEST	0.00	12:55
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:55
BLANK TEST	0.00	12:56

