

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 107308

Date: 08/31/2021

*2AS
 21-34269*

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:09
INTERNAL STANDARD	VERIFIED	11:09
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	11:09
BLANK TEST	0.00	11:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	11:10
BLANK TEST	0.00	11:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	11:11
BLANK TEST	0.00	11:12

