

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 107308

Date: 09/14/2021

RMS
21-36324

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:57
INTERNAL STANDARD	VERIFIED	12:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:58
BLANK TEST	0.00	12:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:59
BLANK TEST	0.00	13:00
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	13:00
BLANK TEST	0.00	13:01

