

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 107308

Date: 09/21/2021

RMS
21-37320

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:50
INTERNAL STANDARD	VERIFIED	12:51
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:51
BLANK TEST	0.00	12:52
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:52
BLANK TEST	0.00	12:53
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:53
BLANK TEST	0.00	12:54

