

**SUPERVISOR TEST**

UTICA CITY  
 POLICE DEPT  
 DATAMASTER dmt: 107308

Date: 09/28/2021

*RMJ*  
*21-38389*

OPERATOR NAME:  
 MICHAEL J CLIVE  
 Permit #: 36682  
 Operator Agency/Dept: UTICA CITY PD  
 Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:15
INTERNAL STANDARD	VERIFIED	12:15
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:15
BLANK TEST	0.00	12:16
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:16
BLANK TEST	0.00	12:17
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:18
BLANK TEST	0.00	12:18

