

SOLUTION CHANGE

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 107308

Date: 03/02/2021

OPERATOR NAME:  
MICHAEL J CLIVE  
Permit #: 36682  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 20500

*RMS*  
*21-7559*

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:09
INTERNAL STANDARD	VERIFIED	13:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	13:09
BLANK TEST	0.00	13:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	13:10
BLANK TEST	0.00	13:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	13:11
BLANK TEST	0.00	13:12

