

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 04/19/2022

TCMS
22-14116

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:31
INTERNAL STANDARD	VERIFIED	13:31
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:31
BLANK TEST	0.00	13:32
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:32
BLANK TEST	0.00	13:33
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:34
BLANK TEST	0.00	13:34

