

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 07/19/2022

*12M >
22-28353*

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:57
INTERNAL STANDARD	VERIFIED	13:57
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:57
BLANK TEST	0.00	13:58
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:58
BLANK TEST	0.00	13:59
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.10	13:59
BLANK TEST	0.00	14:00

