

SUPERVISOR TEST

UTICA CITY  
 POLICE DEPT  
 DATAMASTER dmt: 100282

Date: 07/26/2022

*ZM*  
*22-29518*

OPERATOR NAME:  
 MICHAEL J CLIVE  
 Permit #: 36682  
 Operator Agency/Dept: UTICA CITY PD  
 Reference Standard #: 22100

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 13:17 |
| INTERNAL STANDARD       | VERIFIED | 13:17 |
| REFERENCE STANDARD TEMP | 33.99c   |       |
| REFERENCE STANDARD      | 0.10     | 13:17 |
| BLANK TEST              | 0.00     | 13:18 |
| REFERENCE STANDARD TEMP | 33.99c   |       |
| REFERENCE STANDARD      | 0.10     | 13:18 |
| BLANK TEST              | 0.00     | 13:19 |
| REFERENCE STANDARD TEMP | 33.99c   |       |
| REFERENCE STANDARD      | 0.10     | 13:19 |
| BLANK TEST              | 0.00     | 13:20 |

