

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 100282

Date: 09/06/2022

RM >
22-35906

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:55
INTERNAL STANDARD	VERIFIED	13:55
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:55
BLANK TEST	0.00	13:56
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:57
BLANK TEST	0.00	13:57
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:58
BLANK TEST	0.00	13:59

