

SUPERVISOR TEST

UTICA CITY  
 POLICE DEPT  
 DATAMASTER dmt: 100282

Date: 11/22/2022

*RMS*  
*22-17196*

OPERATOR NAME:  
 MICHAEL J CLIVE  
 Permit #: 36682  
 Operator Agency/Dept: UTICA CITY PD  
 Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:56
INTERNAL STANDARD	VERIFIED	12:57
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:57
BLANK TEST	0.00	12:58
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:58
BLANK TEST	0.00	12:59
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:59
BLANK TEST	0.00	13:00

