

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 107308

ZMS
22-1290

Date: 01/11/2022

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:21
INTERNAL STANDARD	VERIFIED	13:21
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:21
BLANK TEST	0.00	13:22
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:22
BLANK TEST	0.00	13:23
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:23
BLANK TEST	0.00	13:24

