

SUPERVISOR TEST

UTICA CITY  
 POLICE DEPT  
 DATAMASTER dmt: 107308

Date: 05/10/2022

*RMS*  
*22-17209*

OPERATOR NAME:  
 MICHAEL J CLIVE  
 Permit #: 36682  
 Operator Agency/Dept: UTICA CITY PD  
 Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:16
INTERNAL STANDARD	VERIFIED	12:16
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:17
BLANK TEST	0.00	12:17
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:18
BLANK TEST	0.00	12:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:19
BLANK TEST	0.00	12:20

