

SUPERVISOR TEST

UTICA CITY  
 POLICE DEPT  
 DATAMASTER dmt: 107308

Date: 10/04/2022

*RM 5*  
*22-40019*

OPERATOR NAME:  
 MICHAEL J CLIVE  
 Permit #: 36682  
 Operator Agency/Dept: UTICA CITY PD  
 Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:53
INTERNAL STANDARD	VERIFIED	13:53
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	13:54
BLANK TEST	0.00	13:55
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	13:55
BLANK TEST	0.00	13:56
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	13:56
BLANK TEST	0.00	13:57

