

SUPERVISOR TEST

UTICA CITY  
 POLICE DEPT  
 DATAMASTER dmt: 107308

Date: 12/21/2021

*RM S*  
*21-19845*

OPERATOR NAME:  
 MICHAEL J CLIVE  
 Permit #: 36682  
 Operator Agency/Dept: UTICA CITY PD  
 Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:27
INTERNAL STANDARD	VERIFIED	12:27
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:27
BLANK TEST	0.00	12:28
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:28
BLANK TEST	0.00	12:29
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:29
BLANK TEST	0.00	12:30

