

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 100282

Date: 01/24/2023

RMS
23-3668

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:35
INTERNAL STANDARD	VERIFIED	13:35
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:36
BLANK TEST	0.00	13:36
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:37
BLANK TEST	0.00	13:37
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:38
BLANK TEST	0.00	13:39

