

SOLUTION CHANGE

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 100282

Date: 04/03/2023

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 23030

RMS
23-14450

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:52
INTERNAL STANDARD	VERIFIED	14:52
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:53
BLANK TEST	0.00	14:54
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:54
BLANK TEST	0.00	14:55
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:55
BLANK TEST	0.00	14:56

