

SOLUTION CHANGE

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 100282

2M>
23-39985

Date: 09/01/2023

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:05
INTERNAL STANDARD	VERIFIED	13:05
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:05
BLANK TEST	0.00	13:06
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:06
BLANK TEST	0.00	13:07
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:08
BLANK TEST	0.00	13:08

