

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 09/05/2023

EMS
27-10579

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:51
INTERNAL STANDARD	VERIFIED	09:51
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	09:52
BLANK TEST	0.00	09:52
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	09:53
BLANK TEST	0.00	09:53
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	09:54
BLANK TEST	0.00	09:55

