

SOLUTION CHANGE

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 100282

RMS
23-45062

Date: 10/03/2023

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:51
INTERNAL STANDARD	VERIFIED	12:51
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:51
BLANK TEST	0.00	12:52
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:52
BLANK TEST	0.00	12:53
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:53
BLANK TEST	0.00	12:54

