

SUPERVISOR TEST

UTICA CITY  
 POLICE DEPT  
 DATAMASTER dmt: 107308

Date: 01/03/2023

OPERATOR NAME:  
 MICHAEL J CLIVE  
 Permit #: 36682  
 Operator Agency/Dept: UTICA CITY PD  
 Reference Standard #: 22390

*RMS*  
*23-346*

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:39
INTERNAL STANDARD	VERIFIED	11:39
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	11:39
BLANK TEST	0.00	11:40
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	11:40
BLANK TEST	0.00	11:41
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	11:41
BLANK TEST	0.00	11:42

