

SOLUTION CHANGE

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 107308

RM >
23-45062

Date: 10/03/2023

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:57
INTERNAL STANDARD	VERIFIED	14:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:57
BLANK TEST	0.00	14:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:59
BLANK TEST	0.00	14:59
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	15:00
BLANK TEST	0.00	15:00

