

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 107308

Date: 10/10/2023

RMS
23-46144

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:42
INTERNAL STANDARD	VERIFIED	14:42
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:43
BLANK TEST	0.00	14:44
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:44
BLANK TEST	0.00	14:45
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:45
BLANK TEST	0.00	14:46

