

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 107308

Date: 12/05/2023

KMS
23-54567

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:15
INTERNAL STANDARD	VERIFIED	12:15
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:15
BLANK TEST	0.00	12:16
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:16
BLANK TEST	0.00	12:17
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:17
BLANK TEST	0.00	12:18

