

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 107308

Date: 12/12/2023

RMS
23-55397

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:32
INTERNAL STANDARD	VERIFIED	12:32
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	12:33
BLANK TEST	0.00	12:34
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	12:34
BLANK TEST	0.00	12:35
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	12:35
BLANK TEST	0.00	12:36

