

SUPERVISOR TEST

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 100282

Date: 02/27/2024

*RMS*  
*24-8579*

OPERATOR NAME:  
MICHAEL J CLIVE  
Permit #: 36682  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:16
INTERNAL STANDARD	VERIFIED	12:16
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:16
BLANK TEST	0.00	12:17
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:17
BLANK TEST	0.00	12:18
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:18
BLANK TEST	0.00	12:19

