

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 03/14/2024

RM>
24-11304

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:52
INTERNAL STANDARD	VERIFIED	10:52
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	10:52
BLANK TEST	0.00	10:53
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	10:53
BLANK TEST	0.00	10:54
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	10:55
BLANK TEST	0.00	10:55

