

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 03/19/2024

RM
24-12304

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:40
INTERNAL STANDARD	VERIFIED	08:40
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	08:40
BLANK TEST	0.00	08:41
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	08:41
BLANK TEST	0.00	08:42
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	08:43
BLANK TEST	0.00	08:43

