

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 03/26/2024

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 23430

RMS
24-13229

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:36
INTERNAL STANDARD	VERIFIED	09:36
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	09:36
BLANK TEST	0.00	09:37
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	09:37
BLANK TEST	0.00	09:38
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	09:39
BLANK TEST	0.00	09:39

