

SOLUTION CHANGE

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 04/01/2024

LM
24-14231

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:45
INTERNAL STANDARD	VERIFIED	10:45
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	10:46
BLANK TEST	0.00	10:47
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	10:47
BLANK TEST	0.00	10:48
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	10:48
BLANK TEST	0.00	10:49

