

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 04/02/2024

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 24040

KM1
24-14395

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:32
INTERNAL STANDARD	VERIFIED	12:33
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:33
BLANK TEST	0.00	12:34
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:34
BLANK TEST	0.00	12:35
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:35
BLANK TEST	0.00	12:36

