

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 04/09/2024

LMS
21-15808

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:19
INTERNAL STANDARD	VERIFIED	09:19
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	09:19
BLANK TEST	0.00	09:20
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	09:20
BLANK TEST	0.00	09:21
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	09:22
BLANK TEST	0.00	09:22

