

SUPERVISOR TEST

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 100282

Date: 05/07/2024

*AMS*  
*24-20013*

OPERATOR NAME:  
MICHAEL J CLIVE  
Permit #: 36682  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:06
INTERNAL STANDARD	VERIFIED	09:06
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	09:07
BLANK TEST	0.00	09:07
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	09:08
BLANK TEST	0.00	09:09
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	09:09
BLANK TEST	0.00	09:10

